



POLICE DEPARTMENT  
**LAKE HAVASU CITY**  
 2360 McCULLOCH BOULEVARD NORTH  
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281  
 Signature

STATE OF ARIZONA )  
 ) ss  
 COUNTY OF MOHAVE )

Subscribed and sworn to before me this 8 day of April.

2025 by Michael Maloy

Jennifer Lentz  
 NOTARY PUBLIC  
04/08/2025  
 DATE



Commission expires:  
03/15/2029

Administration ..... 855-4884  
 Business Office ..... 855-1171  
 Investigations ..... 855-5775  
 Patrol ..... 855-0515

Area Code 928  
 Fax #'s  
 Administration ..... 680-5430  
 Business Office ..... 680-5431  
 Investigations/Patrol ..... 680-5432

E-mail: police@lhcaz.gov

Emergency ..... 9-1-1  
 Non-Emergency Dispatch/ ..... 855-4111  
 Havasu Silent Witness ..... 854-TIPS

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALBY #281 AGENCY LNCPS  
DATE 04-08-2025 TIME 00:02  
INTOXILYZER SERIAL # 90-003680

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample - Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed

5. Standard Calibration Check:

0.102 AC

- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

PLACED INTO SERVICE AT SIGN & P.S.I. 1000  
FOR THE SUMMER SEASON

SIGNATURE Michael Malby #281

281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

STANDARD CHANGE

Intoxilyzer - Alcohol Analyzer Model 9000

SN: 90-003680 Test Date: 04/07/2025 Test Time: 23:58

Location: -----

Software Version: 9439.01.00 -----

Standard Lot #: 302-402972207 Standard Value: 0.100 Expiration Date: 02/20/2026 Bottle #: 024 Last Changed By: MALLOY, MICHAEL Permit #: 30178

Operator: MALLOY, MICHAEL Permit #:30178 Agency: LAKE HAVASU PD

Test	g/210L	Time
Air Blank	0.000	23:59:06
Calibration Chk	0.101	23:59:28
Air Blank	0.000	00:00:07

No RFI Detected

Successfully Completed Test Sequence

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000 Serial Number: 90-003680 Software Version: 9439.01.00 Analytical Report Number: 36800408250039 OAS: MALLOY, MICHAEL OAS Permit #: 30178 Agency: LAKE HAVASU PD Last 31-Day Check: 12/03/2024 Last Annual Maintenance: 09/25/2024

SUBJECT INFORMATION

Test Date: 04/08/2025 Name: 31-Day Check, Driver's License #: Weight: Gender: State of Issue: Date of Birth: Driver's License Expiration: Age: Deprivation Start Time: 23:45 15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALLOY, MICHAEL Agency: LAKE HAVASU PD Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100 Standard Lot #: 302-402972207 Expiration Date: 02/20/2026 Bottle #: 024 Last Changed By: MALLOY, MICHAEL Permit #: 30178

Test g/210L Time

Air Blank	0.000	00:02:33
Diagnostic Test	PASS	00:03:08
Air Blank	0.000	00:03:45
Calibration Chk	0.102	00:04:07
Air Blank	0.000	00:04:47
Subject Test 1	DEF*	00:08:01
Air Blank	0.000	00:08:45
Air Blank	0.000	00:09:25
Subject Test 2	0.000	00:09:57
Air Blank	0.000	00:10:43
Wait		00:14:46
Air Blank	0.000	00:15:24
Subject Test 3	INJ**	00:15:55
Air Blank	RFI***	00:16:08

\*Deficient Sample \*\*Invalid Sample - Begin new deprivation period \*\*\*RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

\* Deficient Sample \*\* Invalid Sample - Begin new deprivation period \*\*\* RFI Detect All air blank results must be 0.000. Consecutive subject test results must not differ by more than 0.020 g/210L. Standard check results must be ±10%.