



POLICE DEPARTMENT  
**LAKE HAVASU CITY**  
 2360 McCULLOCH BOULEVARD NORTH  
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281  
 Signature

STATE OF ARIZONA        )  
   ) ss  
 COUNTY OF MOHAVE     )


Subscribed and sworn to before me this 3<sup>rd</sup> day of DECEMBER  
2024 by MICHAEL MALOY

[Signature]

NOTARY PUBLIC  
12-3-2024

DATE

Commission expires: 3-16-2025



Area Code 928

Administration ..... 855-4884	Fax #'s	Emergency ..... 9-1-1
Business Office ..... 855-1171	Administration ..... 680-5430	Non-Emergency Dispatch/ ..... 855-4111
Investigations ..... 855-5775	Business Office ..... 680-5431	Havasu Silent Witness ..... 854-TIPS
Patrol ..... 855-0515	Investigations/Patrol ..... 680-5432	

E-mail: police@lhcaz.gov

281

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LUCPD  
DATE 12-03-2024 TIME 01:20  
INTOXILYZER SERIAL # 90-003145

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0. 100 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS: PLACED INTO SERVICE IN JAIL P.S.I: 600  
STANDARD CHANGE

SIGNATURE Michael Maloy #281

281

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003695  
Software Version: 9439.01.00  
Analytical Report Number: 36951203240221

QAS: MALOY, MICHAEL  
QAS Permit #: 30178  
Agency: LAKE HAVASU PD  
Last 31-Day Check: 10/21/2024  
Last Annual Maintenance: 10/28/2024

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 12/03/2024  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 00:30  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MALOY, MICHAEL  
Agency: LAKE HAVASU PD

Permit #: 30178

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 302-402972207  
Expiration Date: 02/20/2026  
Bottle #: 023  
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	01:20:53
Diagnostic Test	PASS	01:21:29
Air Blank	0.000	01:22:06
Calibration Chk	0.100	01:22:27
Air Blank	0.000	01:23:07
Subject Test 1	DEF*	01:26:25
Air Blank	0.000	01:27:06
Air Blank	0.000	01:27:46
Subject Test 2	0.000	01:28:23
Air Blank	0.000	01:29:03
Wait		01:33:08
Air Blank	0.000	01:33:45
Subject Test 3	INV**	01:34:10
Air Blank	RFI***	01:34:23

RESULTS

## EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.

STANDARD CHANGE

281

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003695

Test Date: 12/03/2024

Test Time: 01:18

Location:

Software Version: 9439.01.00

Standard Lot #: 302-402972207

Standard Value: 0.100

Expiration Date: 02/20/2026

Bottle #: 023

Last Changed By: MALOY, MICHAEL

Permit #: 30178

Operator: MALOY, MICHAEL

Permit #: 30178

Agency: LAKE HAVASU PD

Test	g/210L	Time
Air Blank	0.000	01:18:24
Calibration Chk	0.101	01:18:45
Air Blank	0.000	01:19:25

No RFI Detected

Successfully Completed Test Sequence