



POLICE DEPARTMENT
LAKE HAVASU CITY
 2360 McCULLOCH BOULEVARD NORTH
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael Maloy Jr #281
 Signature

STATE OF ARIZONA)
) ss
 COUNTY OF MOHAVE)


Subscribed and sworn to before me this 27 day of February.

2025 by Michael Maloy

Jennifer Le Lentz
 NOTARY PUBLIC

02/27/2025
 DATE

Commission expires:



Area Code 928

Administration 855-4884
 Business Office 855-1171
 Investigations 855-5775
 Patrol 855-0515

Fax #'s
 Administration 680-5430
 Business Office 680-5431
 Investigations/Patrol 680-5432

Emergency 9-1-1
 Non-Emergency Dispatch/ 855-4111
 Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003695
Software Version: 9439.01.00
Analytical Report Number: 36950227250255

QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 01/30/2025
Last Annual Maintenance: 10/28/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 02/27/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 20:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402972207
Expiration Date: 02/20/2026
Bottle #: 022
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	20:27:14
Diagnostic Test	PASS	20:27:50
Air Blank	0.000	20:28:27
Calibration Chk	0.099	20:28:48
Air Blank	0.000	20:29:28
Subject Test 1	DEF*	20:32:46
Air Blank	0.000	20:33:34
Air Blank	0.000	20:34:14
Subject Test 2	0.000	20:35:21
Air Blank	0.000	20:36:11
Wait		20:40:14
Air Blank	0.000	20:40:51
Subject Test 3	INV**	20:41:33
Air Blank	RFI***	20:41:47

RESULTS

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.

281

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LUCPA
DATE 02-27-2025 TIME 20:27
INTOXILYZER SERIAL # 90-003695

- 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 099 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS: STANDARDS CHANGE P.S.I.: 850

SIGNATURE Michael Maloy Jr #281

STANDARDS CHANGE

281

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-003695

Test Date: 02/27/2025

Test Time: 20:23

Location:

Software Version: 9439.01.00

Standard Lot #: 302-402972207

Standard Value: 0.100

Expiration Date: 02/20/2026

Bottle #: 022

Last Changed By: MALOY, MICHAEL

Permit #: 30178

Operator: MALOY, MICHAEL

Permit #: 30178

Agency: LAKE HAVASU PD

Test	g/210L	Time
Air Blank	0.000	20:23:44
Calibration Chk	0.099	20:24:06
Air Blank	0.000	20:24:46

No RFI Detected

Successfully Completed Test Sequence