



EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LNCPD  
DATE 03-27-2025 TIME 17:54  
INTOXILYZER SERIAL # 90-003695

- 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0.098 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

P.S.I.: 600

SIGNATURE Michael Maloy Jr #281

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
 Serial Number: 90-003695  
 Software Version: 9439.01.00  
 Analytical Report Number: 36950327250272

QAS: MALOY, MICHAEL  
 QAS Permit #: 30178  
 Agency: LAKE HAVASU PD  
 Last 31-Day Check: 02/27/2025  
 Last Annual Maintenance: 10/28/2024

## SUBJECT INFORMATION

Name: 31-Day Check  
 Test Date: 03/27/2025  
 Driver's License #:  
 Gender:  
 Date of Birth:  
 Age:

Weight:  
 State of Issue:  
 Driver's License Expiration:  
 Deprivation Start Time: 17:30  
 15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MALOY, MICHAEL  
 Agency: LAKE HAVASU PD

Permit #: 30178

## STANDARD INFORMATION

Standard Value: 0.100  
 Standard Lot #: 302-402972207  
 Expiration Date: 02/20/2026  
 Bottle #: 022  
 Last Changed By: MALOY, MICHAEL

Test	g/210L	Time
Air Blank	0.000	17:54:36
Diagnostic Test	PASS	17:55:12
Air Blank	0.000	17:55:49
Calibration Chk	0.098	17:56:10
Air Blank	0.000	17:56:50
Subject Test 1	DEF*	18:00:08
Air Blank	0.000	18:00:57
Air Blank	0.000	18:01:37
Subject Test 2	0.000	18:02:40
Air Blank	0.000	18:03:23
Wait		18:07:27
Air Blank	0.000	18:08:04
Subject Test 3	INV**	18:08:35
Air Blank	RFI***	18:08:48

Permit #: 30178

**RESULTS**

## EXCEPTION MESSAGES

\* Deficient Sample  
 \*\* Invalid Sample - Begin new deprivation period  
 \*\*\* RFI Detect

## OPERATOR COMMENTS

\*Deficient Sample  
 \*\*Invalid Sample - Begin new deprivation period  
 \*\*\*RFI Detect

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
 Consecutive subject test results must not differ by more than 0.020 g/210L.  
 Standard check results must be ±10%.*