



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.


Michael Maloy Jr #281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 17th day of NOVEMBER.

2024 by MICHAEL MALOY
[Signature]

NOTARY PUBLIC
11-17-2024
DATE



Commission expires: 3-16-2025

Area Code 928

- | | | | | | |
|-----------------------|----------|-----------------------------|----------|-------------------------------|----------|
| Administration | 855-4884 | Fax #'s | | Emergency | 9-1-1 |
| Business Office | 855-1171 | Administration | 680-5430 | Non-Emergency Dispatch/ | 855-4111 |
| Investigations | 855-5775 | Business Office | 680-5431 | Havasu Silent Witness | 854-TIPS |
| Patrol | 855-0515 | Investigations/Patrol | 680-5432 | | |
- E-mail: police@lhcaz.gov

281

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 11-17-2024 TIME 1859
INTOXILYZER SERIAL # 90-0036916

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.099 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

P.S.I. 625

SIGNATURE Michael Maloy Jr #281

ZBI

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003696
Software Version: 9439.01.00
Analytical Report Number: 36961117240077

QAS: MALOY, MICHAEL
QAS Permit #: 30178
Agency: LAKE HAVASU PD
Last 31-Day Check: 10/23/2024
Last Annual Maintenance: 01/10/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 11/17/2024
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 18:40
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402972207
Expiration Date: 02/20/2026
Bottle #: 023
Last Changed By: MALOY, MICHAEL

Permit #: 30178

Test	g/210L	Time
Air Blank	0.000	18:59:14
Diagnostic Test	PASS	18:59:49
Air Blank	0.000	19:00:26
Calibration Chk	0.099	19:00:48
Air Blank	0.000	19:01:29
Subject Test 1	DEF*	19:04:46
Air Blank	0.000	19:05:27
Air Blank	0.000	19:06:07
Subject Test 2	0.000	19:06:42
Air Blank	0.000	19:07:23
Wait		19:11:27
Air Blank	0.000	19:12:04
Subject Test 3	INV**	19:12:44
Air Blank	RFI***	19:12:56

RESULTS

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.