



EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD  
DATE 10-21-2024 TIME 22:35  
INTOXILYZER SERIAL # 90-003695

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0.097 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

SIGN B: TAKEN OUT OF SERVICE FOR THE SEASON AND ANNUAL P.M.

P.S.I.: 875

SIGNATURE Michael Maloy Jr #281

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003695  
Software Version: 9439.01.00  
Analytical Report Number: 36951021240220

QAS: MALOY, MICHAEL  
QAS Permit #: 30178  
Agency: LAKE HAVASU PD  
Last 31-Day Check: 08/27/2024  
Last Annual Maintenance: 11/02/2023

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 10/21/2024  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 22:15  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MALOY, MICHAEL  
Agency: LAKE HAVASU PD

Permit #: 30178

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 302-402801597  
Expiration Date: 03/15/2025  
Bottle #: 001  
Last Changed By: MALOY, MICHAEL

Test	g/210L	Time
Air Blank	0.000	22:35:29
Diagnostic Test	PASS	22:36:05
Air Blank	0.000	22:36:42
Calibration Chk	0.097	22:37:04
Air Blank	0.000	22:37:43
Subject Test 1	DEF*	22:40:58
Air Blank	0.000	22:41:40
Air Blank	0.000	22:42:20
Subject Test 2	0.000	22:42:52
Air Blank	0.000	22:43:35
Wait		22:47:38
Air Blank	0.000	22:48:16
Subject Test 3	INV**	22:48:58
Air Blank	RFI***	22:49:11

Permit #: 30178

## RESULTS

## EXCEPTION MESSAGES

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*