

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LULPO

DATE 02/28/13 TIME 14:59:05

INTOXILYZER SERIAL # 80-001029 LOCATION LULPO JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Earl Wilson

INTOXILYZER 8000

Location: LHCPD MAIN JAIL

Serial Number: 80-001029

Core Version: 8105.48

02/28/2013

14:59:05

Standard Lot#: 22911100A1

Last Changed By: J. SONDRON #187

QAS: E. WILSON #226

LAKE HAVASU CITY PD

Operator: E. WILSON #226

LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	14:59:38
0.100 Cal Check	0.100	14:59:59
Air Blank	0.000	15:00:28

Preventative Maintenance
Performed -- Timer Reset

*EW
2/28*