POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Karl C. E. Dreller is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are a full, true and correct copy of the original.

SQAP SIGNATURE 060409

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 8th day of June, 2009 by KC DRELLER.

Gloria Thomas
NOTARY PUBLIC

6-8-09
DATE

Commission expires: 8-31-10

Notary Public State of Arizona
Mohave County
Gloria Thomas
My Commission Expires
08/31/2010

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@ci.lake-havasu-city.az.us

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K.C. Deller #121 AGENCY Lake Havasu City Police Dept.

DATE 060409 TIME 1527

INTOXILYZER SERIAL # 80-000029 LOCATION LCPD Main

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
   OR
   Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.085 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Deller #121

K.C. Deller #121
Detective

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number: 80-001029
06/04/2009 15:27:09

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:27:37
Cal Check	0.099	15:27:55
Air Blank	0.000	15:28:25

Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K.C. Drexler #121 AGENCY Lake Havasu City Police Dept

DATE 060409 TIME 1528

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD Main

( ) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
( ) 1. Clock time check.
( ) 2. Date check.

OPERATIONAL TESTS
( ) 1. Alcohol-free subject test result 0. 000 AC.
( ) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
( ) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
( ) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES (✓) NO ____

COMMENTS

SIGNATURE K.C. Drexler #121
K.C. Drexler #121
DETECTIVE

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:28:47

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,FREE
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:29:57
Diagnostic Test	Pass	15:30:25
Air Blank	0.000	15:30:52
0.100 Cal Check	0.099	15:31:11
Air Blank	0.000	15:31:40
Subject Test	0.000	15:31:59
Air Blank	0.000	15:32:27
Five Minute Wait	ABT*	
Air Blank	0.000	15:33:02

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:28:47

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,FREE
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:29:57
Diagnostic Test	Pass	15:30:25
Air Blank	0.000	15:30:52
0.100 Cal Check	0.099	15:31:11
Air Blank	0.000	15:31:40
Subject Test	0.000	15:31:59
Air Blank	0.000	15:32:27
Five Minute Wait	ABT*	
Air Blank	0.000	15:33:02

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:28:47

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,FREE
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:29:57
Diagnostic Test	Pass	15:30:25
Air Blank	0.000	15:30:52
0.100 Cal Check	0.099	15:31:11
Air Blank	0.000	15:31:40
Subject Test	0.000	15:31:59
Air Blank	0.000	15:32:27
Five Minute Wait	ABT*	
Air Blank	0.000	15:33:02

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:33:14

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,BAWT
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	15:34:03
Air Blank	0.000	15:34:32

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:33:14

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,BAWT
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	15:34:03
Air Blank	0.000	15:34:32

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:33:14

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,BAWT
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	15:34:03
Air Blank	0.000	15:34:32

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:35:01

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,RFI
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:35:52
Air Blank	0.000	15:36:21

*RFI Detect

Not a Successfully Completed Test Sequence

121

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:35:01

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,RFI
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:35:52
Air Blank	0.000	15:36:21

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:35:01

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,RFI
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:35:52
Air Blank	0.000	15:36:21

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:36:53

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,DEFSAMP
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:38:02
Diagnostic Test	Pass	15:38:29
Air Blank	0.000	15:38:56
0.100 Cal Check	0.099	15:39:15
Air Blank	0.000	15:39:45
Subject Test	0.171*	15:42:53
Air Blank	0.000	15:43:26
Air Blank	0.000	15:43:53
Subject Test	XXX**	15:44:34
Air Blank	0.000	15:45:10

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:36:53

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,DEFSAMP
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:38:02
Diagnostic Test	Pass	15:38:29
Air Blank	0.000	15:38:56
0.100 Cal Check	0.099	15:39:15
Air Blank	0.000	15:39:45
Subject Test	0.171*	15:42:53
Air Blank	0.000	15:43:26
Air Blank	0.000	15:43:53
Subject Test	XXX**	15:44:34
Air Blank	0.000	15:45:10

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
06/04/2009 15:36:53

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,DEFSAMP
DOB: 06/04/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:38:02
Diagnostic Test	Pass	15:38:29
Air Blank	0.000	15:38:56
0.100 Cal Check	0.099	15:39:15
Air Blank	0.000	15:39:45
Subject Test	0.171*	15:42:53
Air Blank	0.000	15:43:26
Air Blank	0.000	15:43:53
Subject Test	XXX**	15:44:34
Air Blank	0.000	15:45:10

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

KCD 121

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K.C. Dealer #121 AGENCY Lake Havasu City Police Dept

DATE 060409 TIME 1547

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD Main

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
   OR
   Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Dealer #121

K.C. Dealer #121
Detective

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number: 80-001029
06/04/2009 15:47:09

Standard Lot#: 833001G
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:47:36
Cal Check	0.098	15:47:55
Air Blank	0.000	15:48:25

K. D. Dreller #121

Operator's Signature