POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 5 day of June, 2015 by Charles Trescher

Laura Varacchi
NOTARY PUBLIC

6/5/2015
DATE

Commission expires:

LAURA VARACCHI
Notary Public - Arizona
Mohave County
My Comm. Expires Feb 27, 2019

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 06-05-15 TIME 1825

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
06/05/2015 18:24:10

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	18:25:03
0.100 Cal Check	0.097	18:25:24
Air Blank	0.000	18:25:53

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 06-05-15 TIME 1826

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
06/05/2015 18:26:45

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:28:18
Diagnostic Test	Pass	18:28:53
Air Blank	0.000	18:29:21
0.100 Cal Check	0.096	18:29:42
Air Blank	0.000	18:30:11
Subject Test	0.000	18:30:28
Air Blank	ABT*	18:30:35
Air Blank	IPS**	18:30:42

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
06/05/2015 18:26:45

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:28:18
Diagnostic Test	Pass	18:28:53
Air Blank	0.000	18:29:21
0.100 Cal Check	0.096	18:29:42
Air Blank	0.000	18:30:11
Subject Test	0.000	18:30:28
Air Blank	ABT*	18:30:35
Air Blank	IPS**	18:30:42

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
06/05/2015 18:26:45

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:28:18
Diagnostic Test	Pass	18:28:53
Air Blank	0.000	18:29:21
0.100 Cal Check	0.096	18:29:42
Air Blank	0.000	18:30:11
Subject Test	0.000	18:30:28
Air Blank	ABT*	18:30:35
Air Blank	IPS**	18:30:42

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
06/05/2015 18:32:38

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:33:38
Diagnostic Test	Pass	18:34:13
Air Blank	0.000	18:34:41
0.100 Cal Check	0.096	18:35:02
Air Blank	0.000	18:35:31
Subject Test	0.058*	18:38:39
Air Blank	0.000	18:39:10
Air Blank	0.000	18:39:38
Subject Test	XXX**	18:40:00
Air Blank	RFI***	18:40:12

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
06/05/2015 18:32:38

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:33:38
Diagnostic Test	Pass	18:34:13
Air Blank	0.000	18:34:41
0.100 Cal Check	0.096	18:35:02
Air Blank	0.000	18:35:31
Subject Test	0.058*	18:38:39
Air Blank	0.000	18:39:10
Air Blank	0.000	18:39:38
Subject Test	XXX**	18:40:00
Air Blank	RFI***	18:40:12

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
06/05/2015 18:32:38

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:33:38
Diagnostic Test	Pass	18:34:13
Air Blank	0.000	18:34:41
0.100 Cal Check	0.096	18:35:02
Air Blank	0.000	18:35:31
Subject Test	0.058*	18:38:39
Air Blank	0.000	18:39:10
Air Blank	0.000	18:39:38
Subject Test	XXX**	18:40:00
Air Blank	RFI***	18:40:12

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 06-05-15 TIME 1842

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.095 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
06/05/2015 18:41:57

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	18:42:29
0.100 Cal Check	0.095	18:42:50
Air Blank	0.000	18:43:19

Preventative Maintenance
Performed -- Timer Reset