POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 7 day of August.

2015 by Josh Sondrol

NOTARY PUBLIC

8/7/15

DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	TDD	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

9/187

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD

DATE 8/7/15 TIME 0217

INTOXILYZER SERIAL # 80-001029 LOCATION LB CHANNEL SIGN 8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

MISSED 31 DAYS DUE TO VACATION

p18 1

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
08/07/2015 02:17:23

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	02:17:54
0.100 Cal Check	0.097	02:18:15
Air Blank	0.000	02:18:44

Preventative Maintenance
Performed -- Timer Reset

4/187

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SONDRAL 187 AGENCY LHC PD

DATE 8/7/15 TIME 0221

INTOXILYZER SERIAL # 80-001029 LOCATION LB CHANNEL SIGN 8

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE J Sondral 187

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
08/07/2015 02:21:47

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	02:22:35
Diagnostic Test	Pass	02:23:10
Air Blank	0.000	02:23:37
0.100 Cal Check	0.097	02:23:58
Air Blank	0.000	02:24:27
Subject Test	0.000	02:24:50
Air Blank	ABT*	02:24:56
Air Blank	IPS**	02:25:04

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

p187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
08/07/2015 02:26:46

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	02:27:32
Diagnostic Test	Pass	02:28:07
Air Blank	0.000	02:28:34
0.100 Cal Check	0.097	02:28:55
Air Blank	0.000	02:29:25
Subject Test	0.241*	02:32:32
Air Blank	0.000	02:33:04
Air Blank	0.000	02:33:32
Subject Test	XXX**	02:33:58
Air Blank	RFI***	02:34:10

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDRA 187 AGENCY LHC PD

DATE 8/7/15 TIME 0235

INTOXILYZER SERIAL # 80-001029 LOCATION LB CHANNEL SIGN 8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondra 187

DPS Form Exh G-2 (Rev 05-01)

7187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
08/07/2015 02:35:04

Standard Lot#: 01314100A3
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	02:35:33
0.100 Cal Check	0.098	02:35:54
Air Blank	0.000	02:36:23

Preventative Maintenance
Performed -- Timer Reset