POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Earl Wilson is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 17 day of AUGUST.

2012 by CARLA DUARTE

NOTARY PUBLIC

Aug. 17, 2012

DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	TDD	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LAKE HAVASU CITY P.D.

DATE 08/17/12 TIME 10:31:23

INTOXILYZER SERIAL # 80-003355 LOCATION BRIDGEWATER CHANNEL SIGN #8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE EL Wilson 226

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
08/17/2012 10:31:23

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	10:32:01
0.100 Cal Check	0.098	10:32:22
Air Blank	0.000	10:32:52

Preventative Maintenance
Performed -- Timer Reset

EW-226

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LAKE HAVASU CITY P.D.

DATE 08/17/12 TIME 10:38 09

INTOXILYZER SERIAL # 80-003355 LOCATION BRIDGEWATER CHANNEL SIGN #8

(✓ 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓ 1. Clock time check.
(✓ 2. Date check.

OPERATIONAL TESTS
(✓ 1. Alcohol-free subject test result 0. 000 AC.
(✓ 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓ 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓ 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Earl Wilson 226

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
08/17/2012 10:38:09

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 12/12/2010
Sex: M Weight: 110

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:39:42
Diagnostic Test	Pass	10:40:17
Air Blank	0.000	10:40:44
0.100 Cal Check	0.097	10:41:05
Air Blank	0.000	10:41:35
Subject Test	0.000	10:41:51
Air Blank	ABT*	10:42:02
Air Blank	IPS**	10:42:11

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

EW-226

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
08/17/2012 10:43:34

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 12/12/2010
Sex: M Weight: 110

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:44:42
Diagnostic Test	Pass	10:45:17
Air Blank	0.000	10:45:44
0.100 Cal Check	0.098	10:46:05
Air Blank	0.000	10:46:35
Subject Test	0.144*	10:49:42
Air Blank	0.000	10:50:17
Air Blank	0.000	10:50:44
Subject Test	XXX**	10:51:06
Air Blank	RFI***	10:51:14

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EMW-226

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LAKE HAVASU CITY P.D.

DATE 08/17/12 TIME 10:57:27

INTOXILYZER SERIAL # 80 003355 LOCATION BRIDGEWATER CHANNEL SIGN #8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.096 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE E L L 226

DPS Form Exh G-2 (Rev 05-01)

CONDUCTED G2 @ 10:55:37 WHICH REVEALED A "AMBIENT FAIL".

SECOND G2 CONDUCTED @ 10:57:27 COMPLETED CALIBRATION WITH NO ERRORS.

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80 003355
Core Version: 8105.48
08/17/2012 10:55:37

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	AMB*	10:56:10
Air Blank	0.000	10:56:38

*Ambient Fail

EW 226

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80 003355
Core Version: 8105.48
08/17/2012 10:57:27

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	10:57:59
0.100 Cal Check	0.096	10:58:20
Air Blank	0.000	10:58:50

Preventative Maintenance
Performed -- Timer Reset

EW 226