POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 2nd day of June, 2016 by Charles Trescher

NOTARY PUBLIC

6-2-16
DATE

Commission expires: 2-26-19

LORI SCHNEPF
Notary Public - Arizona
Mohave County
My Comm. Expires Feb 26, 2019

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

CT
#219

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 6-2-16 TIME 1348 HRS

INTOXILYZER SERIAL # 80-001873 LOCATION LHCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [signature]

DPS Form Exh G-2 (Rev 05-01)

CT
#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
06/02/2016 13:46:41

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	13:47:31
0.100 Cal Check	0.098	13:47:51
Air Blank	0.000	13:48:21

Preventative Maintenance
Performed -- Timer Reset

CT
#219

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST CHARLES TRESCNER AGENCY LAKE HAVASU CITY POLICE

DATE 6-2-16 TIME 1401 HRS

INTOXILYZER SERIAL # 80-001873 LOCATION LHCPD JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC. Result 0.099 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE

DPS Form Exh G-4 (Rev 05-01)

CT
#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
06/02/2016 13:49:31

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 215

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:50:35
Diagnostic Test	Pass	13:51:10
Air Blank	0.000	13:51:38
0.100 Cal Check	0.097	13:51:59
Air Blank	0.000	13:52:29
Subject Test	0.000	13:52:56
Air Blank	ABT*	13:53:03
Air Blank	IPS**	13:53:10

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

CT
#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
06/02/2016 13:49:31

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 215

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:50:35
Diagnostic Test	Pass	13:51:10
Air Blank	0.000	13:51:38
0.100 Cal Check	0.097	13:51:59
Air Blank	0.000	13:52:29
Subject Test	0.000	13:52:56
Air Blank	ABT*	13:53:03
Air Blank	IPS**	13:53:10

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

CT
#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
06/02/2016 13:49:31

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 215

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:50:35
Diagnostic Test	Pass	13:51:10
Air Blank	0.000	13:51:38
0.100 Cal Check	0.097	13:51:59
Air Blank	0.000	13:52:29
Subject Test	0.000	13:52:56
Air Blank	ABT*	13:53:03
Air Blank	IPS**	13:53:10

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

CT
#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
06/02/2016 13:54:19

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TES
DOB: 10/19/1978
Sex: M Weight: 215

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:55:16
Diagnostic Test	Pass	13:55:51
Air Blank	0.000	13:56:18
0.100 Cal Check	0.099	13:56:39
Air Blank	0.000	13:57:09
Subject Test	0.031*	14:00:16
Air Blank	0.000	14:00:47
Air Blank	0.000	14:01:15
Subject Test	XXX**	14:01:32
Air Blank	RFI***	14:01:39

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

CT
*219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
06/02/2016 13:54:19

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TES
DOB: 10/19/1978
Sex: M Weight: 215

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:55:16
Diagnostic Test	Pass	13:55:51
Air Blank	0.000	13:56:18
0.100 Cal Check	0.099	13:56:39
Air Blank	0.000	13:57:09
Subject Test	0.031*	14:00:16
Air Blank	0.000	14:00:47
Air Blank	0.000	14:01:15
Subject Test	XXX**	14:01:32
Air Blank	RFI***	14:01:39

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

CT
#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
06/02/2016 13:54:19

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TES
DOB: 10/19/1978
Sex: M Weight: 215

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:55:16
Diagnostic Test	Pass	13:55:51
Air Blank	0.000	13:56:18
0.100 Cal Check	0.099	13:56:39
Air Blank	0.000	13:57:09
Subject Test	0.031*	14:00:16
Air Blank	0.000	14:00:47
Air Blank	0.000	14:01:15
Subject Test	XXX**	14:01:32
Air Blank	RFI***	14:01:39

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

CT
#219

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCNER AGENCY LAKE HAVASU CITY POLICE

DATE 6-2-16 TIME

INTOXILYZER SERIAL # 80-60 1873 LOCATION LHCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

DPS Form Exh G-2 (Rev 05-01)

CT
#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
06/02/2016 14:03:05

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	AMB*	14:03:37
Air Blank	0.000	14:04:05

*Ambient Fail

CT
#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
06/02/2016 14:04:09

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	14:04:39
0.100 Cal Check	0.099	14:05:00
Air Blank	0.000	14:05:29

Preventative Maintenance
Performed -- Timer Reset