

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LHC PD
DATE 11/29/16 TIME 0014
INTOXILYZER SERIAL # 80-001873 LOCATION MAIN JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

J Sondrol 187

INTOXILYZER 8000
Location: LEC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
11/29/2016

00:14:10

Standard Lot#: 2231510CA1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	00:14:45
0.100 Cal Check	0.097	00:15:06
Air Blank	0.000	00:15:36

Preventative Maintenance
Performed -- Timer Reset

187

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SONNDEL 187 AGENCY LNC PD
DATE 11/29/16 TIME 0017
INTOXILYZER SERIAL # 80-001873 LOCATION MARIN JAIL

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE J Sonndel 187

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
11/29/2016

00:17:56

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:18:44
Diagnostic Test	Pass	00:19:19
Air Blank	0.000	00:19:46
0.100 Cal Check	0.097	00:20:07
Air Blank	0.000	00:20:37
Subject Test	0.000	00:20:55
Air Blank	ABT*	00:21:02
Air Blank	IPS**	00:21:09

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
11/29/2016

00:22:25

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP, TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:23:26
Diagnostic Test	Pass	00:24:01
Air Blank	0.000	00:24:28
0.100 Cal Check	0.097	00:24:49
Air Blank	0.000	00:25:19
Subject Test	0.056*	00:28:26
Air Blank	0.000	00:28:58
Air Blank	0.000	00:29:26
Subject Test	XXX**	00:29:54
Air Blank	RFI***	00:30:19

*Deficient Sample
**Invalid Sample
 Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SANDRIL 187 AGENCY LNC PD
DATE 11/29/16 TIME 0032
INTOXILYZER SERIAL # 80-001873 LOCATION MAIN JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0.097 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sandril 187

INTOXILYZER 8000

Location: LHC MAIN JAIL

Serial Number: 80-001873

Core Version: 8105.48

11/29/2016

00:32:05

Standard Lot#: 22315100A1

Last Changed By: J. SONDRON #187

QAS: J. SONDRON #187

LAKE HAVASU CITY PD

Operator: J. SONDRON #187

LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	00:32:34
0.100 Cal Check	0.097	00:32:55
Air Blank	0.000	00:33:25

Preventative Maintenance
Performed -- Timer Reset